



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994

KONICA MINOLTA SENSING AMERICAS, INC.
101 Williams Drive
Ramsey, NJ 07446
Perry Fisher, Mark Lombardi Phone: 201 236 4300

CALIBRATION

Valid To: January 31, 2026

Certificate Number: 4673.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1,4}:

I. Optical Quantities

Parameter/Equipment	Range	CMC ^{2,3} (±)	Comments
Illuminance – Illuminance Meters	(8, 80, 800) lux	1.9 %	Illuminance meter master body (calibrated reference instrument)
Luminance – Luminance Meters	(135 to 165) cd/m ²	2.2 %	Luminance meter master body (calibrated reference instrument)
Correlated Color Temperature (CCT) – Lamp	(2836 to 2876) K	22 K	Colorimeter master body (calibrated reference instrument)
Correlated Color Temperature (CCT) – Illuminance Meters	(2836 to 2876) K	22 K	CCT standard (lamp)
Chromaticity (x, y) – Illuminance Meters	$0.4457 \leq x \leq 0.4488$ $0.4069 \leq y \leq 0.4078$	$\Delta x \leq 0.0024$ $\Delta y \leq 0.0023$	CCT standard (lamp)

Parameter/Equipment	Range	CMC ^{2,3} (±)	Comments
Chromaticity (x, y) – Luminance Meters	$0.4457 \leq x \leq 0.4488$ $0.4069 \leq y \leq 0.4078$	$\Delta x \leq 0.0024$ $\Delta y \leq 0.0023$	CCT standard (lamp) and radiance coefficient standard (diffuser)
di:8° (SCI) Spectral Reflectance Factor	(80 to 100) % 360 nm 370 nm (400 to 470) nm (480 to 720) nm 730 nm 740 nm	1.4 % 1.3 % 0.76 % 0.71 % 0.75 % 0.81 %	Reflectance standard tile
de:8° (SCE) Spectral Reflectance Factor	(80 to 100) % 360 nm 370 nm (400 to 470) nm (480 to 720) nm 730 nm 740 nm	1.5 % 1.3 % 0.78 % 0.74 % 0.77 % 0.83 %	Reflectance standard tile
45°:0° Spectral Reflectance Factor	(80 to 100) % (360 to 370) nm (380 to 390) nm (400 to 740) nm	1.4 % 1.3 % 0.9 %	Reflectance standard tile
Illuminance – Spectroradiometer	(130 to 1200) lux (350 to 1100) nm	2.1 %	Incandescent standard lamp, end rod
Luminance – Spectroradiometer	(450 to 550) cd/m ² (360 to 1100) nm	2.2 %	Integrating sphere light source
Luminous Flux – Spectroradiometer	0.05 mlm to 250 klm (360 to 1100) nm	2.1 %	Standard lamp, integrating sphere
Luminous Flux for LEDs in Conformity with CIE 127:2007	(0.5 to 3.5) lm (360 to 1100) nm	2.5 %	LED luminous flux standard, integrating sphere
Averaged LED Intensity: ILED-B in Conformity with CIE 127:2007	(0.25 to 1.0) cd (360 to 1100) nm	2.5 %	LED luminous intensity standard

Parameter/Equipment	Range	CMC ^{2,3} (±)	Comments
Spectral Irradiance – [Spectroradiometer] (370 to 1100) nm	(370 to 390) nm (400 to 430) nm (440 to 990) nm (1000 to 1050) nm (1060 to 1100) nm	4.0 % 2.7 % 2.4 % 4.0 % 4.6 %	Spectral intensity standard (lamp), distance standard (end rod)
Spectral Radiance – [Spectroradiometer] (380 to 1040) nm	(380 to 390) nm (400 to 540) nm (550 to 1040) nm	4.7 % 3.6 % 3.1 %	Spectral radiance standard (integrating sphere light source)

¹ This laboratory offers commercial calibration service for Konica Minolta and Instrument Systems equipment only.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.

³ In the statement of CMC, percentages are percentage of reading, unless otherwise indicated.

⁴ This scope meets A2LA's *P112 Flexible Scope Policy*.



Accredited Laboratory

A2LA has accredited

KONICA MINOLTA SENSING AMERICAS, INC.

Ramsey, NJ

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCCL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 26th day of January 2024.

A blue ink signature of Mr. Trace McInturff, written in a cursive style.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 4673.01
Valid to January 31, 2026

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.