



KONICA MINOLTA

Spectrophotometer

CM-5

An advanced all-in-one spectrophotometer with innovative operation to let anyone take measurements easily anytime



Mini Petri Dish
(optional accessory) enables
measurements of small sample
amounts.
• Automatic white calibration
compatible.

The CM-5 makes color measurements simple. Just switch it on and start taking measurements. No need to bother with a computer; the CM-5 has a full range of advanced functions including specialized indices for a variety of applications and a large color display that makes results easy to read.

Finally, high accuracy and ease of use in a compact top-port spectrophotometer!

Main Specifications

Model	Spectrophotometer CM-5	
Illumination/viewing system	Reflectance:	di:8°, de:8° (diffuse illumination: 8° viewing) SCI (specular component included)/SCE (specular component excluded) switchable Conforms to CIE No. 15 (2004), ISO 7724/1, ASTM E 1164, DIN 5033 Teil 7, and JIS Z 8722 condition c standard.
	Transmittance:	di:0°, de:0° (diffuse illumination: 0° viewing) Conforms to CIE No. 15 (2004), ASTM E 1164, DIN 5033 Teil 7, and JIS Z 8722 condition g standard.
Integrating sphere size	Ø152 mm	
Detector	Dual 40-element silicon photodiode arrays	
Spectral separation device	Planar diffraction grating	
Wavelength range	360 nm to 740 nm	
Wavelength pitch	10 nm	
Half bandwidth	Approx. 10 nm	
Measurement range	0 to 175 % (Reflectance or transmittance); Output/display resolution: 0.01%	
Light source	Pulsed xenon lamp	
Measurement time	Approx. 1 s (to data display/output); Minimum measurement interval: Approx. 3 s	
Measurement/Illumination area	Reflectance:	Changeable by changing mask and settings. LAV: Ø30 mm/Ø36 mm; MAV (optional): Ø8 mm/Ø11 mm; SAV (optional): Ø3 mm/Ø6 mm
	Transmittance:	Approx. Ø20 mm/ Ø26 mm
Repeatability	Spectral reflectance: Standard deviation within 0.1% (400 nm to 740 nm) Chromaticity value: Standard deviation within ΔE*ab 0.04 (When a white calibration plate is measured 30 times at 10-second intervals after white calibration)	
Inter-instrument agreement	Within ΔE*ab 0.15 (Typical) (LAV/SCI) (Based on 12 BCRA Series II color tiles compared to values measured with a master body under Konica Minolta standard conditions)	
Transmittance chamber	No sides (unlimited sample length); Depth (maximum sample thickness): 60 mm Sample holders (optional) for holding sheet samples or containers of liquid samples can be installed/removed	
Display	5.7-inch TFT color LCD	
Display languages	English, Japanese, German, French, Italian, Spanish, Simplified Chinese, Portuguese	
White/100% calibration	Automatic white (reflectance)/100% (transmittance) calibration using internal white calibration plate (Not applicable to 100% calibration when using cells for transmittance measurements of liquids.)	
Interfaces	USB 1.1(Connection to PC, USB memory *1, USB keyboard *2); RS-232C standard (Connection to serial printer)	
Observer	2° or 10° Standard Observer	
Illuminant	A, C, D50, D65, F2, F6, F7, F8, F10, F11, F12, ID50, ID65 (simultaneous evaluation with two light sources possible)	
Displayed data	Spectral values/graph (reflectance, transmittance, absorbance), colorimetric values/graph, color-difference values/graph, pass/fail judgment, pseudocolor, color assessment	
Colorimetric data	L*a*b*, L*C*h, Hunter Lab, XyY, XYZ, and color differences in these spaces; Munsell	
Indexes	Reflectance:	MI; WI (ASTM E 313-73, ASTM E 313-96, Hunter); YI (ASTM E 313-73, ASTM E 313-96, ASTM D 1925); WB (ASTM E 313-73); ISO Brightness
	Transmittance:	Gardner; Iodine Color Number; Platinum-Cobalt Color Scale (Hazen/ APHA); European Pharmacopoeia; US Pharmacopeia
User index	User-defined index (Optional CM-CT1(Ver.1.4 or later) and Color Management Software SpectraMagic™ NX2 required for setting user index.) * User index can be used with CM-5/CR-5 firmware version 1.2 or later.	
Color-difference equation	ΔE*ab (CIE 1976), ΔE*94 (CIE 1994), ΔE00 (CIE DE2000), ΔE (Hunter), CMC (l:c)	
Pass/Fail judgment	Tolerances can be set to colorimetric values (except Munsell), color-difference values, or reflectance index values	
Data memory	Measurement data: 4,000 measurements; Target color data: 1,000 measurements	
USB memory*1 storage	Storage of measurement data and target color data. Storage/reading of measurement condition settings	
Power	AC 100 to 240 V, 50/60 Hz (using exclusive AC Adapter)	
Size (W × H × D)	Slide cover closed: 385 × 192 × 261mm Slide cover open: 475 × 192 × 261 mm	
Weight	Approx. 5.8 kg	
Operation temperature/ humidity range	13 to 33°C, relative humidity 80 % or less (at 33°C) with no condensation	
Storage temperature/ humidity range	0 to 40°C, relative humidity 80 % or less (at 35°C) with no condensation	

*1 USB memory devices with no security features are supported.

*2 USB human interface device class US layout keyboards are supported. (Operation is not guaranteed for all of the above supported USB memories and keyboards.)



SAFETY PRECAUTIONS

For correct use and for your safety, be sure to read the instruction manual before using the instrument.

- Always connect the instrument to the specified power supply voltage. Improper connection may cause a fire or electric shock.

• The specifications and appearance shown herein are subject to change without notice.

• If you have any questions about specifications, please contact your Konica Minolta representative.

• Konica Minolta, the Konica Minolta logo and symbol mark, and "Giving Shape to Ideas" are registered trademarks or trademarks of KONICA MINOLTA, INC.

ISO Certifications of KONICA MINOLTA, Inc., Sakai Site



JQA-QMA15988
Design, development, manufacture/
manufacturing management, calibration, and
service of measuring instruments



JQA-E-80027
Design, development, manufacture, service and sales
of measuring instruments

KONICA MINOLTA, INC.

Osaka, Japan
New Jersey, U.S.A. PHONE: (888)473-2656 (in USA), +1(201)236-4300 (outside USA) FAX: +1(201)785-2480 E-Mail: marketing.sus@konicaminolta.com

Konica Minolta Sensing Americas, Inc.

European HQ/ BENELUX Nieuwegein, Netherlands PHONE: +31(0)30 248-1193 E-Mail: info.benelux@seu.konicaminolta.eu

Konica Minolta Sensing Europe B.V.

German Office München, Germany PHONE: +49(0)89 4357 156 0 E-Mail: info.germany@seu.konicaminolta.eu
French Office Roissy CDG Cedex, France PHONE: +33(0)1 80 11 10 70 E-Mail: info.france@seu.konicaminolta.eu
UK Office Warrington, United Kingdom PHONE: +44(0)1925 467300 E-Mail: info.uk@seu.konicaminolta.eu
Italian Office Ciniselli Balsamo, Italy PHONE: +39 02849488.00 E-Mail: info.italy@seu.konicaminolta.eu
Swiss Office Dietikon, Switzerland PHONE: +41(0)43 322-9800 E-Mail: info.switzerland@seu.konicaminolta.eu
Nordic Office VÄSTRA FRÖLUNDA, Sweden PHONE: +46(0)31 709946 E-Mail: info.nordic@seu.konicaminolta.eu
Polish Office Wrocław, Poland PHONE: +48(0)71 73452-11 E-Mail: info.poland@seu.konicaminolta.eu

Konica Minolta (CHINA) Investment Ltd.

SE Sales Division Shanghai, China PHONE: +86(0)21-6057-1089 E-Mail: hcn_sensing@gcp.konicaminolta.com
Beijing Office Beijing, China PHONE: +86(0)10-8522 1551 E-Mail: hcn_sensing@gcp.konicaminolta.com
Guangzhou Office Guangzhou, China PHONE: +86(0)20-3826 4220 E-Mail: hcn_sensing@gcp.konicaminolta.com
Chongqing Office Chongqing, China PHONE: +86(0)23-6773 4988 E-Mail: hcn_sensing@gcp.konicaminolta.com

Qingdao Office Shandong, China PHONE: +86(0)532-8079 1871 E-Mail: hcn_sensing@gcp.konicaminolta.com
Wuhan Office Hubei, China PHONE: +86(0)27-6885 0586 E-Mail: hcn_sensing@gcp.konicaminolta.com
Shenzhen Office Shenzhen, China PHONE: +86(0)755-2868 7535 E-Mail: hcn_sensing@gcp.konicaminolta.com
Xiamen Office Xiamen, China PHONE: +86(0)592-7107 399 E-Mail: hcn_sensing@gcp.konicaminolta.com

Konica Minolta Sensing Singapore Pte. Ltd.

Singapore PHONE: +65 6563-5533 E-Mail: ssg@gcp.konicaminolta.com

Konica Minolta Sensing Korea Co., Ltd.

Korean HQ Goyang-si, Korea PHONE: +82(0)2-523-9726 E-Mail: se.korea@konicaminolta.com
Cheonan Office Cheonan-si, Korea PHONE: +82(0)41-556-9726 E-Mail: se.korea@konicaminolta.com

Addresses and telephone/fax numbers and e-mail address are subject to change without notice.
For the latest contact information, please refer to KONICA MINOLTA Worldwide Offices web page.

<https://konicaminolta.com/instruments/network>